

KEY NOTES

1. VERIFY WITH USER THE USE OF F0265, CHAIR, VERSES M4665, STRETCHER.
2. PROVIDE A1165, IV TRACKS, FOR NAVY PROJECTS ONLY.
3. PROVIDE AT LEAST ONE HANDWASHING FACILITY OR ONE FOR EVERY THIRD RECOVERY MODULE. P3200, ELECTRONIC LAVATORY, MAY BE USED INSTEAD OF P3100, LAVATORY. FOR THE P3200 PROVIDE EITHER BATTERY POWER OR HARD WIRED CONNECTION.
4. SIZE AND CONFIGURATION OF E0406, NURSE STATION, MAY VARY.
5. LENGTH OF E1500, WALL MOUNTED RAIL, TO BE DETERMINED BY MANUFACTURER/INSTALLER.
6. SEE ELEVATION OF A1107, UTILITY RAIL SYSTEM, FOR UTILITIES.
7. ALL CABINETS SHALL BE LOCKING.

RECOVERY CUBICLE, PHASE 2 (RROP1)

4.04.06

SURGERY (INPATIENT & AMBULATORY)


0 1000 2000

FLOOR/EQUIPMENT PLAN



DATE: 01/28/2002

1:50 SCALE (MILLIMETERS)



KEY NOTES

1. PROVIDE A1165, IV TRACKS, FOR NAVY PROJECTS ONLY.
2. LIGHT FIXTURE IS CONTROLLED BY SWITCH LOCATED IN A1110, PREFABRICATED HEADWALL. SEE HEADWALL ELEVATION FOR SWITCH ELEVATION.

RECOVERY CUBICLE, PHASE 2 (RROP1)

4.04.06

SURGERY (INPATIENT & AMBULATORY)


0 1000 2000

CEILING PLAN




DATE: 01/28/2002

1:50 SCALE (MILLIMETERS)



WEST



NORTH

EAST

NOTES

1. SEE TYPICAL EQUIPMENT MOUNTING HEIGHTS BEGINNING ON SHEET 7.01.01.

KEY NOTES

1. PROVIDE A1165, IV TRACKS, FOR NAVY PROJECTS ONLY.
2. PROVIDE AT LEAST ONE HANDWASHING FACILITY OR ONE FOR EVERY THIRD RECOVERY MODULE. P3200, ELECTRONIC LAVATORY, MAY BE USED INSTEAD OF P3100, LAVATORY. FOR THE P3200 PROVIDE EITHER BATTERY POWER OR HARD WIRED CONNECTION.
3. LENGTH OF E1500, WALL MOUNTED RAIL, TO BE DETERMINED BY MANUFACTURER/INSTALLER.
4. ALL CABINETS SHALL BE LOCKING.

RECOVERY CUBICLE, PHASE 2 (RROP1)

4.04.06

SURGERY (INPATIENT & AMBULATORY)


0 1000 2000

ELEVATIONS



DATE: 01/28/2002

1:50 SCALE (MILLIMETERS)



NOTES

1. THE UTILITIES AND DEVICES INDICATED ARE FOR AN EMERGENCY ROOM IN A HOSPITAL.
2. ALL RECEPTACLES IN HEADWALL SHALL BE HOSPITAL GRADE.

KEY NOTES

1. PROVIDE A1165, IV TRACKS, FOR NAVY PROJECTS ONLY.
2. VACUUM SLIDE.

RECOVERY CUBICLE, PHASE 2 (RROP1)

4.04.06

SURGERY (INPATIENT & AMBULATORY)

0 500 1000

FLOOR/EQUIPMENT PLAN



DATE: 01/28/2002

1:30 SCALE (MILLIMETERS)

Medical Equipment Room Guide Plates

4.04 Surgery (Inpatient & Ambulatory)

Plate Number 4.04.06
Room Name RECOVERY CUBICLE, PHASE 2
SEPS # RROP1

JSN	Nomenclature	Qty	Army	Navy	AF	Utilities					
			Cat	Cat	Cat	1	2	3	4	5	6
A1107	Rail System , Utility, Gas and Electric	2	A	A	A	.	F	G	.	.	A
A1165	Track, IV, Ceiling Mounted, 7 Foot	4	A	A	A	J
A5075	Dispenser, Soap, Disposable	2	C	C	C
A5080	Dispenser, Paper Towel, SS, Surface Mounted	2	A	A	A	J
A5106	Waste Disposal Unit, Sharps w/Glove Dispenser	3	C	A	C
A5180	Track, Cubicle, Surface Mounted, With Curtain	2	A	A	A	J
C03P0	Cabinet, Sink, U/C/B, 2 Door, 36x30x22	1	A	A	A
C04H0	Cabinet, U/C/B, 2 Half Drawer, 3 Drawer, 36x36x22	1	A	A	A
CD040	Cabinet, W/H, 2 Shelf, 2 DO, Sloping Top, 38x36x13	1	A	A	A	J
CS010	Sink, SS, Single Compartment, 7.5x12x12 ID	1	A	A	A	D
CT030	Countertop, High Pressure Laminate	6	A	A	A
E0406	Nurse Station, Angle, Free Standing, 96" W	1	F	F	F	.	A
E0906	Locker, Supply, General, Wall Mtd, 23"W x 20"D	1	C	C	C
E0912	Locker, Supply, Med Surg, Wall Mtd, 23"W x 20"D	1	C	C	C
E0915	Locker, Supply, Medication, Wall Mtd, 23"W x 20"D	1	C	C	C
E0948	Cart, General Storage, Mobile, 42"H x 32"W x 22"D	2	C	C	C
E0954	Cart, Emergency, Mobile, 66"H x 52"W x 22"D	1	C	C	C
E1500	Rail, MOD, W/MNTD, HX144XD	1	C	C	C	J
F0205	Chair, Side With Arms	2	C	C	C
F0265	Chair, Recliner	1	C	C	C
F0280	Chair, Swivel, Low Back	1	C	C	C
F2000	Basket, Wastepaper, Round, Metal, 18 H x 16 Dia.	5	C	C	C
F3200	Clock, Battery, 12" Diameter	2	C	C	C
M0750	Flowmeter, Air, Connect w/50 PSI Supply	2	C	C	C	.	.	K	.	.	.
M0755	Flowmeter, Oxygen, Low Flow	2	C	C	C	.	.	A	.	.	.
M0765	Regulator, Vacuum	6	C	C	C	.	.	B	.	.	.
M1800	Computer, Microprocessing	1	C	C	C	.	A
M1825	Printer, Computer	1	C	C	C	.	A
M1855	Facsimile Machine	1	C	C	C	.	A
M3070	Hamper, Linen, Mobile, w/Lid	1	C	C	C
M3072	Frame, Infectious Waste Bag w/Lid	1	C	C	C
M3105	Cabinet, Warming, F/S, 1 Heated Compartment, Elect	1	A	A	A	.	F
M3150	Distribution System, Medication, Automatic	1	C	C	C	.	A
M4100	Sphygmomanometer, Aneroid, Wall Mounted	2	C	C	C
M4116	Monitor, Vital Signs	3	C	C	C	.	A
M4255	Stand, IV, Adjustable	2	C	C	C
M4266	Pump, Volumetric, Infusion, Multiple Line	2	C	C	C	.	A
M4665	Stretcher, Recovery, Surgical	1	C	C	C
M7435	Light, Overbed, Direct And Indirect	2	A	A	A	.	F	.	.	.	J

Medical Equipment Room Guide Plates

M7655	Defibrillator/Cardioscope, 5-Lead	1	C	C	C	.	A
M7845	Monitor, Physiological, Bedside, 4 Channel	2	C	C	C	.	A	.	.	.	D
M7850	Monitor, Physiological, Central, 8 Bed, Color	1	C	C	C	.	A	.	.	.	D
M7905	Oximeter, Pulse	2	C	C	C	.	A
P3100	Lavatory, Vitreous China, Slab Type	1	A	A	A	D